

Government Engineering College Barton Hill,  
Thiruvananthapuram  
Email. p1@gecbh.ac.in

No. GECBH/179/2026-P1

Dated:17-04-2026

NOTICE

**Quotation No: 18/2025-26**

Sealed quotations are invited for the **Purchase of Visible Spectrophotometer for the Environmental Engineering Lab of Civil Engineering Department of this Institution** specified in the schedule attached below/overleaf. The rates quoted should be for delivery of the articles at the place mentioned below the schedule. The necessary superscription, due date for the receipt of quotations, the date up to which the rates will have to remain firm for acceptance and the name and address of officer to whom the quotation is to be sent are noted below. Any quotation received after the time fixed on the due date is liable to be rejected. The maximum period required for delivery of the articles should also be mentioned. Quotations not stipulating period of firmness and with price variation clause and/or 'subject to prior sale' condition are liable to be rejected.

The prices quoted should be inclusive of all taxes, duties, GST etc. which are or may become payable by the contractor under existing or future law or rules of the country of origin/supply or delivery during the course of execution of the contract.

Special conditions, if any, printed on the quotation sheets of the tenders or attached with the tender will not be applicable to the contract unless they are expressly accepted in writing by the purchaser.

**Superscription : Purchase of Spectrophotometer for the Environmental Engineering Lab of Civil Engineering Department of this Institution.**

**Quotation No: 18/2025-26**

**Due date and time for receipt of quotations: 05/05/2026 2pm**

**Date and Time for opening Quotation: 05/05/2026 3pm**

**Date up to which the rates are to remain firm for acceptance: 04/04/2027**

**Designation and address of Officer whom the quotation is to be addressed:  
Principal Government Engineering College, Barton Hill**

Place: Thiruvananthapuram

Date:17-04-2026



**Bijulal D  
Principal(i/c)**

Detailed Specifications attached below:

Copy to

1. Sri Rajesh Kumar K, System Analyst (for publishing on the college website)
2. Smt.Monisha.V, Assistant Professor in Civil Engg.Department.1
3. Notice Board

**SPECIFICATION OF VISIBLE SPECTROPHOTOMETER FOR THE ENVIRONMENTAL ENGG. LAB OF CIVIL ENGG. DEPARTMENT**

<b>Sl.No</b>	<b>Item</b>	<b>Details</b>
<b>1</b>	<b>Parameter</b>	<b>Required Specification</b>
<b>2</b>	<b>Wavelength Range</b>	<b>340 nm 960nm (Visible Range)</b>
<b>3</b>	<b>Optical System</b>	<b>Single Beam with 1200 lines / mm grating</b>
<b>4</b>	<b>Accuracy</b>	<b>±0.005 Abs</b>
<b>5</b>	<b>Sample holder</b>	<b>4 positions, 10 mm to 50 mm</b>



**Bijulal D  
PRINCIPAL**